

December 07, 2020

Chad Hess  
Poland Central School District  
74 Cold Brook Street  
Poland, NY 13431

RE: Project: leads 11//20  
Pace Project No.: 70154624

Dear Chad Hess:

Enclosed are the analytical results for sample(s) received by the laboratory on November 24, 2020. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Julie Litvin  
julie.litvin@pacelabs.com  
(631)694-3040  
Project Manager

Enclosures

cc: Laura Dutton, Poland Central School District  
Adam Hutchinson, Poland Central School District



## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## CERTIFICATIONS

Project: leads 11//20

Pace Project No.: 70154624

---

### **Pace Analytical Services Long Island**

575 Broad Hollow Rd, Melville, NY 11747

New York Certification #: 10478 Primary Accrediting Body

New Jersey Certification #: NY158

Pennsylvania Certification #: 68-00350

Connecticut Certification #: PH-0435

Maryland Certification #: 208

Rhode Island Certification #: LAO00340

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

---

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 1</b>		<b>Lab ID: 70154624001</b>	Collected: 11/20/20 06:55	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		12/04/20 14:46	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 2</b>		<b>Lab ID: 70154624002</b>	Collected: 11/20/20 07:05	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>3.4</b>	ug/L	1.0	1		12/04/20 14:48	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 4</b>		<b>Lab ID: 70154624003</b>	Collected: 11/20/20 06:59	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 14:49	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 5		Lab ID: 70154624004		Collected: 11/20/20 07:00		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		12/04/20 14:51	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 6</b>		<b>Lab ID: 70154624005</b>	Collected: 11/20/20 07:01	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 14:56	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 7</b>		<b>Lab ID: 70154624006</b>	Collected: 11/20/20 06:56	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 15:03	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 8</b>		<b>Lab ID: 70154624007</b>	Collected: 11/20/20 06:57	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 15:05	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 9		Lab ID: 70154624008		Collected: 11/20/20 06:58		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.0	ug/L	1.0	1		12/04/20 15:07	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 10</b>		<b>Lab ID: 70154624009</b>	Collected: 11/20/20 07:02	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>28.1</b>	ug/L	1.0	1		12/04/20 15:08	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 11</b>		<b>Lab ID: 70154624010</b>	Collected: 11/20/20 07:47	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		12/04/20 15:10	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 12</b>		<b>Lab ID: 70154624011</b>	Collected: 11/20/20 07:11	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>127</b>	ug/L	1.0	1		12/04/20 15:11	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 13 (SINK 2)</b>		<b>Lab ID: 70154624012</b>	Collected: 11/20/20 07:13	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>157</b>	ug/L	1.0	1		12/04/20 15:13	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 15		Lab ID: 70154624013		Collected: 11/20/20 06:49		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	132	ug/L	1.0	1		12/04/20 15:15	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 19		Lab ID: 70154624014		Collected: 11/20/20 06:46		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:20	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 20		Lab ID: 70154624015		Collected: 11/20/20 07:09		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	18.2	ug/L	1.0	1		12/04/20 15:21	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 21		Lab ID: 70154624016		Collected: 11/20/20 06:48		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	11.6	ug/L	1.0	1		12/04/20 15:26	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 22		Lab ID: 70154624017		Collected: 11/20/20 06:43		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.6	ug/L	1.0	1		12/04/20 15:28	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 23		Lab ID: 70154624018		Collected: 11/20/20 06:41		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	4.3	ug/L	1.0	1		12/04/20 15:29	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 24		Lab ID: 70154624019		Collected: 11/20/20 05:42		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:31	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 25		Lab ID: 70154624020		Collected: 11/20/20 05:41		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.8	ug/L	1.0	1		12/04/20 15:32	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 26</b>		<b>Lab ID: 70154624021</b>	Collected: 11/20/20 05:19	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		12/04/20 15:34	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 28		Lab ID: 70154624022		Collected: 11/20/20 05:20		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:39	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 30		Lab ID: 70154624023		Collected: 11/20/20 05:22		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:41	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 32		Lab ID: 70154624024		Collected: 11/20/20 05:25		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:42	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 33		Lab ID: 70154624025		Collected: 11/20/20 05:26		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:47	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 34		Lab ID: 70154624026		Collected: 11/20/20 05:30		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.6	ug/L	1.0	1		12/04/20 15:52	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 35		Lab ID: 70154624027		Collected: 11/20/20 05:31		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:53	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 36		Lab ID: 70154624028		Collected: 11/20/20 05:33		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 15:58	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 37		Lab ID: 70154624029		Collected: 11/20/20 05:34		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 16:00	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 38		Lab ID: 70154624030		Collected: 11/20/20 05:37		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.0	ug/L	1.0	1		12/04/20 16:01	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 40</b>		<b>Lab ID: 70154624031</b>	Collected: 11/20/20 05:36	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		12/04/20 16:03	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 42</b>		<b>Lab ID: 70154624032</b>	Collected: 11/20/20 05:32	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 16:05	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 44</b>		<b>Lab ID: 70154624033</b>	Collected: 11/20/20 05:27	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 16:06	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 45</b>		<b>Lab ID: 70154624034</b>	Collected: 11/20/20 05:28	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>1.8</b>	ug/L	1.0	1		12/04/20 16:08	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 46		Lab ID: 70154624035		Collected: 11/20/20 05:23		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 16:10	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 48		Lab ID: 70154624036		Collected: 11/20/20 05:21		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	5.9	ug/L	1.0	1		12/04/20 16:18	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 50</b>		<b>Lab ID: 70154624037</b>	Collected: 11/20/20 05:17	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 16:19	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 52		Lab ID: 70154624038		Collected: 11/20/20 05:15		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		12/04/20 16:21	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 54		Lab ID: 70154624039		Collected: 11/20/20 05:13		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		12/04/20 16:22	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 55		Lab ID: 70154624040		Collected: 11/20/20 05:14		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.6	ug/L	1.0	1		12/04/20 16:24	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 56</b>		<b>Lab ID: 70154624041</b>	Collected: 11/20/20 05:12	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 16:26	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 58 BOYS</b>		<b>Lab ID: 70154624042</b>	Collected: 11/20/20 05:52	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>4.8</b>	ug/L	1.0	1		12/04/20 16:27	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 60 GIRLS		Lab ID: 70154624043		Collected: 11/20/20 05:53		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.9	ug/L	1.0	1		12/04/20 16:29	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 61 GIRLS</b>		<b>Lab ID: 70154624044</b>	Collected: 11/20/20 05:56	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>4.8</b>	ug/L	1.0	1		12/04/20 16:31	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 63		Lab ID: 70154624045		Collected: 11/20/20 05:47		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.8	ug/L	1.0	1		12/04/20 16:39	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 64		Lab ID: 70154624046		Collected: 11/20/20 05:48		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	5.6	ug/L	1.0	1		12/04/20 16:43	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 65		Lab ID: 70154624047		Collected: 11/20/20 05:58		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	6.8	ug/L	1.0	1		12/04/20 16:45	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 66		Lab ID: 70154624048		Collected: 11/20/20 06:06		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	7.2	ug/L	1.0	1		12/04/20 16:47	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 67</b>		<b>Lab ID: 70154624049</b>	Collected: 11/20/20 06:07	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>6.4</b>	ug/L	1.0	1		12/04/20 16:48	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 68		Lab ID: 70154624050		Collected: 11/20/20 06:08		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 16:50	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 69</b>		<b>Lab ID: 70154624051</b>	Collected: 11/20/20 06:01	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>5.3</b>	ug/L	1.0	1		12/04/20 16:52	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 70</b>		<b>Lab ID: 70154624052</b>	Collected: 11/20/20 06:02	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>3.5</b>	ug/L	1.0	1		12/04/20 16:56	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 71		Lab ID: 70154624053		Collected: 11/20/20 06:03		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	8.7	ug/L	1.0	1		12/04/20 16:58	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 72</b>		<b>Lab ID: 70154624054</b>	Collected: 11/20/20 06:04	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>106</b>	ug/L	1.0	1		12/04/20 17:00	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 73		Lab ID: 70154624055		Collected: 11/20/20 06:14		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.8	ug/L	1.0	1		12/04/20 17:01	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 74		Lab ID: 70154624056		Collected: 11/20/20 06:15		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		12/04/20 17:06	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 75		Lab ID: 70154624057		Collected: 11/20/20 06:16		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.0	ug/L	1.0	1		12/04/20 17:08	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 76		Lab ID: 70154624058		Collected: 11/20/20 06:17		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	5.1	ug/L	1.0	1		12/04/20 17:09	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 77</b>		<b>Lab ID: 70154624059</b>	Collected: 11/20/20 06:19	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.1	ug/L	1.0	1		12/04/20 17:11	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 78		Lab ID: 70154624060		Collected: 11/20/20 06:20		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.3	ug/L	1.0	1		12/04/20 17:16	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 79</b>		<b>Lab ID: 70154624061</b>	Collected: 11/20/20 06:22	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 17:17	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 80</b>		<b>Lab ID: 70154624062</b>	Collected: 11/20/20 06:23	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 17:19	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 81		Lab ID: 70154624063		Collected: 11/20/20 06:36		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.5	ug/L	1.0	1		12/04/20 17:21	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 82		Lab ID: 70154624064		Collected: 11/20/20 06:37		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.1	ug/L	1.0	1		12/04/20 17:22	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 83</b>		<b>Lab ID: 70154624065</b>	Collected: 11/20/20 06:38	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 17:27	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 84		Lab ID: 70154624066		Collected: 11/20/20 06:39		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.1	ug/L	1.0	1		12/04/20 17:35	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 85</b>		<b>Lab ID: 70154624067</b>	Collected: 11/20/20 07:16	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 17:37	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 86</b>		<b>Lab ID: 70154624068</b>	Collected: 11/20/20 08:37	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>1.0</b>	ug/L	1.0	1		12/04/20 17:38	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 87</b>		<b>Lab ID: 70154624069</b>	Collected: 11/20/20 07:18	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>4.4</b>	ug/L	1.0	1		12/04/20 17:40	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 88		Lab ID: 70154624070		Collected: 11/20/20 07:19		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 17:42	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 89</b>		<b>Lab ID: 70154624071</b>	Collected: 11/20/20 07:21	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.4</b>	ug/L	1.0	1		12/04/20 17:43	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 90</b>		<b>Lab ID: 70154624072</b>	Collected: 11/20/20 07:22	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>6.5</b>	ug/L	1.0	1		12/07/20 12:07	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 91		Lab ID: 70154624073		Collected: 11/20/20 07:23		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	8.5	ug/L	1.0	1		12/04/20 17:46	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 92</b>		<b>Lab ID: 70154624074</b>	Collected: 11/20/20 07:24	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		12/04/20 17:48	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 93		Lab ID: 70154624075		Collected: 11/20/20 07:26		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 17:50	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 94</b>		<b>Lab ID: 70154624076</b>	Collected: 11/20/20 07:27	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 17:58	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 95		Lab ID: 70154624077		Collected: 11/20/20 07:29		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/04/20 17:59	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 96		Lab ID: 70154624078		Collected: 11/20/20 07:30		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.6	ug/L	1.0	1		12/04/20 18:01	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 97		Lab ID: 70154624079		Collected: 11/20/20 07:31		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.1	ug/L	1.0	1		12/04/20 18:03	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 98</b>		<b>Lab ID: 70154624080</b>	Collected: 11/20/20 07:38	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:04	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 99</b>		<b>Lab ID: 70154624081</b>	Collected: 11/20/20 07:39	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:06	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 100</b>		<b>Lab ID: 70154624082</b>	Collected: 11/20/20 07:40	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:07	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 101</b>		<b>Lab ID: 70154624083</b>	Collected: 11/20/20 07:41	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:09	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 102</b>		<b>Lab ID: 70154624084</b>	Collected: 11/20/20 07:42	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:14	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 103</b>		<b>Lab ID: 70154624085</b>	Collected: 11/20/20 07:44	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:19	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 104</b>		<b>Lab ID: 70154624086</b>	Collected: 11/20/20 07:45	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:24	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 107		Lab ID: 70154624087		Collected: 11/20/20 07:49		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	26.4	ug/L	1.0	1		12/04/20 18:25	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 108		Lab ID: 70154624088		Collected: 11/20/20 07:50		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	26.7	ug/L	1.0	1		12/04/20 18:27	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 109</b>		<b>Lab ID: 70154624089</b>	Collected: 11/20/20 07:51	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>39.0</b>	ug/L	1.0	1		12/04/20 18:29	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 110</b>		<b>Lab ID: 70154624090</b>	Collected: 11/20/20 07:52	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>34.3</b>	ug/L	1.0	1		12/04/20 18:33	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 116</b>		<b>Lab ID: 70154624091</b>	Collected: 11/20/20 07:36	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.8</b>	ug/L	1.0	1		12/04/20 18:35	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 117</b>		<b>Lab ID: 70154624092</b>	Collected: 11/20/20 08:09	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:37	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 118		Lab ID: 70154624093		Collected: 11/20/20 08:05		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	9.5	ug/L	1.0	1		12/04/20 18:38	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 119</b>		<b>Lab ID: 70154624094</b>	Collected: 11/20/20 08:06	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>4.2</b>	ug/L	1.0	1		12/04/20 18:40	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 120		Lab ID: 70154624095		Collected: 11/20/20 08:07		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.8	ug/L	1.0	1		12/04/20 18:42	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 121		Lab ID: 70154624096		Collected: 11/20/20 08:08		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		12/04/20 18:46	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 122</b>		<b>Lab ID: 70154624097</b>	Collected: 11/20/20 08:00	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.9</b>	ug/L	1.0	1		12/04/20 18:48	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 123</b>		<b>Lab ID: 70154624098</b>	Collected: 11/20/20 07:57	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>4.8</b>	ug/L	1.0	1		12/04/20 18:53	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 125		Lab ID: 70154624099		Collected: 11/20/20 07:59		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	16.5	ug/L	1.0	1		12/04/20 18:54	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 127</b>		<b>Lab ID: 70154624100</b>	Collected: 11/20/20 08:03	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 18:56	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 128</b>		<b>Lab ID: 70154624101</b>	Collected: 11/20/20 07:55	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>10.4</b>	ug/L	1.0	1		12/04/20 18:58	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 130		Lab ID: 70154624102		Collected: 11/20/20 08:11		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	39.3	ug/L	1.0	1		12/04/20 18:59	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 131		Lab ID: 70154624103		Collected: 11/20/20 08:12		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	39.0	ug/L	1.0	1		12/04/20 19:01	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 132</b>		<b>Lab ID: 70154624104</b>	Collected: 11/20/20 08:13	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>59.2</b>	ug/L	1.0	1		12/04/20 19:03	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 133</b>		<b>Lab ID: 70154624105</b>	Collected: 11/20/20 08:14	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>31.9</b>	ug/L	1.0	1		12/04/20 19:07	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 134</b>		<b>Lab ID: 70154624106</b>	Collected: 11/20/20 08:15	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>24.9</b>	ug/L	1.0	1		12/04/20 19:15	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 135</b>		<b>Lab ID: 70154624107</b>	Collected: 11/20/20 08:16	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>57.4</b>	ug/L	1.0	1		12/04/20 19:17	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 136		Lab ID: 70154624108		Collected: 11/20/20 08:17		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	47.7	ug/L	1.0	1		12/04/20 19:19	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 137</b>		<b>Lab ID: 70154624109</b>	Collected: 11/20/20 08:18	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>54.9</b>	ug/L	1.0	1		12/04/20 19:20	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 138		Lab ID: 70154624110		Collected: 11/20/20 08:21		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	16.5	ug/L	1.0	1		12/04/20 19:22	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 139</b>		<b>Lab ID: 70154624111</b>	Collected: 11/20/20 08:22	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>34.6</b>	ug/L	1.0	1		12/04/20 19:24	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 140</b>		<b>Lab ID: 70154624112</b>	Collected: 11/20/20 08:23	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>34.9</b>	ug/L	1.0	1		12/04/20 19:25	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 141		Lab ID: 70154624113		Collected: 11/20/20 08:24		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	115	ug/L	1.0	1		12/04/20 19:27	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 142</b>		<b>Lab ID: 70154624114</b>	Collected: 11/20/20 08:26	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>20.3</b>	ug/L	1.0	1		12/04/20 19:32	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 143</b>		<b>Lab ID: 70154624115</b>	Collected: 11/20/20 08:27	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>25.5</b>	ug/L	1.0	1		12/04/20 19:33	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 144		Lab ID: 70154624116		Collected: 11/20/20 08:28		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	24.9	ug/L	1.0	1		12/04/20 19:38	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 145</b>		<b>Lab ID: 70154624117</b>	Collected: 11/20/20 08:29	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>76.8</b>	ug/L	1.0	1		12/04/20 19:40	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 147</b>		<b>Lab ID: 70154624118</b>	Collected: 11/20/20 08:31	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		12/04/20 19:41	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 150		Lab ID: 70154624119		Collected: 11/20/20 06:11		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.6	ug/L	1.0	1		12/04/20 19:43	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 151		Lab ID: 70154624120		Collected: 11/20/20 06:12		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.2	ug/L	1.0	1		12/04/20 19:45	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: <b>PCSD152 CAFE CLASS SINK 1 LEFT</b>		Lab ID: <b>70154624121</b>	Collected: 11/20/20 06:28	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>6.3</b>	ug/L	1.0	1		12/04/20 19:46	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 153 - 2</b>		<b>Lab ID: 70154624122</b>	Collected: 11/20/20 06:29	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.4</b>	ug/L	1.0	1		12/04/20 19:51	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 154 - 3</b>		<b>Lab ID: 70154624123</b>	Collected: 11/20/20 06:30	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.3</b>	ug/L	1.0	1		12/04/20 19:53	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 155 - 4</b>		<b>Lab ID: 70154624124</b>	Collected: 11/20/20 06:31	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>5.6</b>	ug/L	1.0	1		12/04/20 19:54	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

Sample: PCSD 156 - 5		Lab ID: 70154624125		Collected: 11/20/20 06:32		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	11.7	ug/L	1.0	1		12/04/20 19:59	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 157 CLASS SINK 6 RIGHT</b>		<b>Lab ID: 70154624126</b>	Collected: 11/22/20 06:33	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>104</b>	ug/L	1.0	1		12/04/20 20:04	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 158 RM 62 CLASS SINK</b>		<b>Lab ID: 70154624127</b>	Collected: 11/22/20 06:42	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<b>2.0</b>	ug/L	1.0	1		12/04/20 20:05	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 159 ACROSS 56 LOW DF BOT</b>		<b>Lab ID: 70154624128</b>		Collected: 11/22/20 07:25		Received: 11/24/20 10:45		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<b>&lt;1.0</b>	ug/L	1.0	1		12/04/20 20:10	7439-92-1		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: leads 11//20

Pace Project No.: 70154624

<b>Sample: PCSD 160 2ND FLOOR STAFF BATH</b>		<b>Lab ID: 70154624129</b>	Collected: 11/22/20 08:02	Received: 11/24/20 10:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/04/20 20:12	7439-92-1	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch: 188255

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70154624001, 70154624002, 70154624003, 70154624004

METHOD BLANK: 923416

Matrix: Water

Associated Lab Samples: 70154624001, 70154624002, 70154624003, 70154624004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 14:02	

LABORATORY CONTROL SAMPLE: 923417

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.8	104	85-115	

MATRIX SPIKE SAMPLE: 923420

Parameter	Units	70155238001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	4	4.7	114	70-130	

MATRIX SPIKE SAMPLE: 923422

Parameter	Units	70154888019 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	3.4	4	11.2	114	70-130	

SAMPLE DUPLICATE: 923419

Parameter	Units	70155238001 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 923421

Parameter	Units	70154888019 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	3.4	6.7	1	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch:	188257	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624005, 70154624006, 70154624007, 70154624008, 70154624009, 70154624010, 70154624011, 70154624012, 70154624013, 70154624014, 70154624015, 70154624016, 70154624017, 70154624018, 70154624019, 70154624020, 70154624021, 70154624022, 70154624023, 70154624024		

METHOD BLANK:	923424	Matrix:	Water
Associated Lab Samples:	70154624005, 70154624006, 70154624007, 70154624008, 70154624009, 70154624010, 70154624011, 70154624012, 70154624013, 70154624014, 70154624015, 70154624016, 70154624017, 70154624018, 70154624019, 70154624020, 70154624021, 70154624022, 70154624023, 70154624024		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 14:52	

LABORATORY CONTROL SAMPLE:	923425					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	52.2	104	85-115	

MATRIX SPIKE SAMPLE:		923427					
		70154624005	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	4	5.5	112	70-130	

MATRIX SPIKE SAMPLE:		923429					
Parameter	Units	70154624015 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	18.2	4	22.7	113	70-130	

SAMPLE DUPLICATE:	923426				
Parameter	Units	70154624005 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	923428				
Parameter	Units	70154624015 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	18.2	18.1	0	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20  
Pace Project No.: 70154624

QC Batch:	188258	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624025, 70154624026, 70154624027, 70154624028, 70154624029, 70154624030, 70154624031, 70154624032, 70154624033, 70154624034, 70154624035, 70154624036, 70154624037, 70154624038, 70154624039, 70154624040, 70154624041, 70154624042, 70154624043, 70154624044		

METHOD BLANK:	923430	Matrix:	Water
Associated Lab Samples:	70154624025, 70154624026, 70154624027, 70154624028, 70154624029, 70154624030, 70154624031, 70154624032, 70154624033, 70154624034, 70154624035, 70154624036, 70154624037, 70154624038, 70154624039, 70154624040, 70154624041, 70154624042, 70154624043, 70154624044		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 15:44	

LABORATORY CONTROL SAMPLE:	923431					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.8	102	85-115	

MATRIX SPIKE SAMPLE:		923433					
		70154624025	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	4	5.7	128	70-130	

MATRIX SPIKE SAMPLE:		923435					
		70154624035	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	4	5.9	129	70-130	

SAMPLE DUPLICATE:	923432				
Parameter	Units	70154624025 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	923434				
Parameter	Units	70154624035 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch:	188259	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624045, 70154624046, 70154624047, 70154624048, 70154624049, 70154624050, 70154624051, 70154624052, 70154624053, 70154624054, 70154624055, 70154624056, 70154624057, 70154624058, 70154624059, 70154624060, 70154624061, 70154624062, 70154624063, 70154624064		

METHOD BLANK:	923436	Matrix:	Water
Associated Lab Samples:	70154624045, 70154624046, 70154624047, 70154624048, 70154624049, 70154624050, 70154624051, 70154624052, 70154624053, 70154624054, 70154624055, 70154624056, 70154624057, 70154624058, 70154624059, 70154624060, 70154624061, 70154624062, 70154624063, 70154624064		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 16:32	

LABORATORY CONTROL SAMPLE:	923437					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE:		923439					
		70154624045	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	3.8	4	8.8	124	70-130	

MATRIX SPIKE SAMPLE:		923441					
		70154624055	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	1.8	4	6.5	119	70-130	

SAMPLE DUPLICATE: 923438					
		70154624045	Dup		
Parameter	Units	Result	Result	RPD	Qualifiers
Lead	ug/L	3.8	3.8	2	

SAMPLE DUPLICATE: 923440

Parameter	Units	70154624055 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.8	1.8	1	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch:	188260	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624065, 70154624066, 70154624067, 70154624068, 70154624069, 70154624070, 70154624071, 70154624072, 70154624073, 70154624074, 70154624075, 70154624076, 70154624077, 70154624078, 70154624079, 70154624080, 70154624081, 70154624082, 70154624083, 70154624084		

METHOD BLANK: 923442

Matrix: Water

Associated Lab Samples: 70154624065, 70154624066, 70154624067, 70154624068, 70154624069, 70154624070, 70154624071, 70154624072, 70154624073, 70154624074, 70154624075, 70154624076, 70154624077, 70154624078, 70154624079, 70154624080, 70154624081, 70154624082, 70154624083, 70154624084

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 17:24	

LABORATORY CONTROL SAMPLE: 923443

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.6	101	85-115	

MATRIX SPIKE SAMPLE: 923445

Parameter	Units	70154624065 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	4	5.5	113	70-130	

MATRIX SPIKE SAMPLE: 923447

Parameter	Units	70154624075 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	4	4.6	112	70-130	

SAMPLE DUPLICATE: 923444

Parameter	Units	70154624065 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 923446

Parameter	Units	70154624075 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch:	188261	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624085, 70154624086, 70154624087, 70154624088, 70154624089, 70154624090, 70154624091, 70154624092, 70154624093, 70154624094, 70154624095, 70154624096, 70154624097, 70154624098, 70154624099, 70154624100, 70154624101, 70154624102, 70154624103, 70154624104		

METHOD BLANK:	923448	Matrix:	Water
Associated Lab Samples:	70154624085, 70154624086, 70154624087, 70154624088, 70154624089, 70154624090, 70154624091, 70154624092, 70154624093, 70154624094, 70154624095, 70154624096, 70154624097, 70154624098, 70154624099, 70154624100, 70154624101, 70154624102, 70154624103, 70154624104		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 18:16	

LABORATORY CONTROL SAMPLE:	923449					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE:		923451					
		70154624085	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	4	5.3	112	70-130	

MATRIX SPIKE SAMPLE:		923453					
		70154624095	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	1.8	4	6.2	109	70-130	

SAMPLE DUPLICATE:	923450				
Parameter	Units	70154624085 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	923452				
Parameter	Units	70154624095 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.8	1.8	3	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20  
Pace Project No.: 70154624

QC Batch:	188262	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70154624105, 70154624106, 70154624107, 70154624108, 70154624109, 70154624110, 70154624111, 70154624112, 70154624113, 70154624114, 70154624115, 70154624116, 70154624117, 70154624118, 70154624119, 70154624120, 70154624121, 70154624122, 70154624123, 70154624124		

METHOD BLANK:	923454	Matrix:	Water
Associated Lab Samples:	70154624105, 70154624106, 70154624107, 70154624108, 70154624109, 70154624110, 70154624111, 70154624112, 70154624113, 70154624114, 70154624115, 70154624116, 70154624117, 70154624118, 70154624119, 70154624120, 70154624121, 70154624122, 70154624123, 70154624124		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 19:04	

LABORATORY CONTROL SAMPLE:	923455					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.9	102	85-115	

MATRIX SPIKE SAMPLE:		923457					
		70154624105	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	31.9	4	35.5	90	70-130	

MATRIX SPIKE SAMPLE:		923459					
		70154624115	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	25.5	4	30.0	114	70-130	

SAMPLE DUPLICATE:	923456				
Parameter	Units	70154624105 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	31.9	31.1	3	

SAMPLE DUPLICATE:	923458				
Parameter	Units	70154624115 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	25.5	25.6	0	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA

Project: leads 11//20

Pace Project No.: 70154624

QC Batch: 188263

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70154624125, 70154624126, 70154624127, 70154624128, 70154624129

METHOD BLANK: 923460

Matrix: Water

Associated Lab Samples: 70154624125, 70154624126, 70154624127, 70154624128, 70154624129

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/04/20 19:56	

LABORATORY CONTROL SAMPLE: 923461

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.0	100	85-115	

MATRIX SPIKE SAMPLE: 923463

Parameter	Units	70154624125 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	11.7	4	16.4	118	70-130	

SAMPLE DUPLICATE: 923462

Parameter	Units	70154624125 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	11.7	11.7	0	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALIFIERS

Project: leads 11//20

Pace Project No.: 70154624

---

### DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: leads 11//20

Pace Project No.: 70154624

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70154624001	PCSD 1	EPA 200.8	188255		
70154624002	PCSD 2	EPA 200.8	188255		
70154624003	PCSD 4	EPA 200.8	188255		
70154624004	PCSD 5	EPA 200.8	188255		
70154624005	PCSD 6	EPA 200.8	188257		
70154624006	PCSD 7	EPA 200.8	188257		
70154624007	PCSD 8	EPA 200.8	188257		
70154624008	PCSD 9	EPA 200.8	188257		
70154624009	PCSD 10	EPA 200.8	188257		
70154624010	PCSD 11	EPA 200.8	188257		
70154624011	PCSD 12	EPA 200.8	188257		
70154624012	PCSD 13 (SINK 2)	EPA 200.8	188257		
70154624013	PCSD 15	EPA 200.8	188257		
70154624014	PCSD 19	EPA 200.8	188257		
70154624015	PCSD 20	EPA 200.8	188257		
70154624016	PCSD 21	EPA 200.8	188257		
70154624017	PCSD 22	EPA 200.8	188257		
70154624018	PCSD 23	EPA 200.8	188257		
70154624019	PCSD 24	EPA 200.8	188257		
70154624020	PCSD 25	EPA 200.8	188257		
70154624021	PCSD 26	EPA 200.8	188257		
70154624022	PCSD 28	EPA 200.8	188257		
70154624023	PCSD 30	EPA 200.8	188257		
70154624024	PCSD 32	EPA 200.8	188257		
70154624025	PCSD 33	EPA 200.8	188258		
70154624026	PCSD 34	EPA 200.8	188258		
70154624027	PCSD 35	EPA 200.8	188258		
70154624028	PCSD 36	EPA 200.8	188258		
70154624029	PCSD 37	EPA 200.8	188258		
70154624030	PCSD 38	EPA 200.8	188258		
70154624031	PCSD 40	EPA 200.8	188258		
70154624032	PCSD 42	EPA 200.8	188258		
70154624033	PCSD 44	EPA 200.8	188258		
70154624034	PCSD 45	EPA 200.8	188258		
70154624035	PCSD 46	EPA 200.8	188258		
70154624036	PCSD 48	EPA 200.8	188258		
70154624037	PCSD 50	EPA 200.8	188258		
70154624038	PCSD 52	EPA 200.8	188258		
70154624039	PCSD 54	EPA 200.8	188258		
70154624040	PCSD 55	EPA 200.8	188258		
70154624041	PCSD 56	EPA 200.8	188258		
70154624042	PCSD 58 BOYS	EPA 200.8	188258		
70154624043	PCSD 60 GIRLS	EPA 200.8	188258		
70154624044	PCSD 61 GIRLS	EPA 200.8	188258		
70154624045	PCSD 63	EPA 200.8	188259		
70154624046	PCSD 64	EPA 200.8	188259		
70154624047	PCSD 65	EPA 200.8	188259		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: leads 11//20

Pace Project No.: 70154624

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70154624048	PCSD 66	EPA 200.8	188259		
70154624049	PCSD 67	EPA 200.8	188259		
70154624050	PCSD 68	EPA 200.8	188259		
70154624051	PCSD 69	EPA 200.8	188259		
70154624052	PCSD 70	EPA 200.8	188259		
70154624053	PCSD 71	EPA 200.8	188259		
70154624054	PCSD 72	EPA 200.8	188259		
70154624055	PCSD 73	EPA 200.8	188259		
70154624056	PCSD 74	EPA 200.8	188259		
70154624057	PCSD 75	EPA 200.8	188259		
70154624058	PCSD 76	EPA 200.8	188259		
70154624059	PCSD 77	EPA 200.8	188259		
70154624060	PCSD 78	EPA 200.8	188259		
70154624061	PCSD 79	EPA 200.8	188259		
70154624062	PCSD 80	EPA 200.8	188259		
70154624063	PCSD 81	EPA 200.8	188259		
70154624064	PCSD 82	EPA 200.8	188259		
70154624065	PCSD 83	EPA 200.8	188260		
70154624066	PCSD 84	EPA 200.8	188260		
70154624067	PCSD 85	EPA 200.8	188260		
70154624068	PCSD 86	EPA 200.8	188260		
70154624069	PCSD 87	EPA 200.8	188260		
70154624070	PCSD 88	EPA 200.8	188260		
70154624071	PCSD 89	EPA 200.8	188260		
70154624072	PCSD 90	EPA 200.8	188260		
70154624073	PCSD 91	EPA 200.8	188260		
70154624074	PCSD 92	EPA 200.8	188260		
70154624075	PCSD 93	EPA 200.8	188260		
70154624076	PCSD 94	EPA 200.8	188260		
70154624077	PCSD 95	EPA 200.8	188260		
70154624078	PCSD 96	EPA 200.8	188260		
70154624079	PCSD 97	EPA 200.8	188260		
70154624080	PCSD 98	EPA 200.8	188260		
70154624081	PCSD 99	EPA 200.8	188260		
70154624082	PCSD 100	EPA 200.8	188260		
70154624083	PCSD 101	EPA 200.8	188260		
70154624084	PCSD 102	EPA 200.8	188260		
70154624085	PCSD 103	EPA 200.8	188261		
70154624086	PCSD 104	EPA 200.8	188261		
70154624087	PCSD 107	EPA 200.8	188261		
70154624088	PCSD 108	EPA 200.8	188261		
70154624089	PCSD 109	EPA 200.8	188261		
70154624090	PCSD 110	EPA 200.8	188261		
70154624091	PCSD 116	EPA 200.8	188261		
70154624092	PCSD 117	EPA 200.8	188261		
70154624093	PCSD 118	EPA 200.8	188261		
70154624094	PCSD 119	EPA 200.8	188261		
70154624095	PCSD 120	EPA 200.8	188261		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: leads 11//20

Pace Project No.: 70154624

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70154624096	PCSD 121	EPA 200.8	188261		
70154624097	PCSD 122	EPA 200.8	188261		
70154624098	PCSD 123	EPA 200.8	188261		
70154624099	PCSD 125	EPA 200.8	188261		
70154624100	PCSD 127	EPA 200.8	188261		
70154624101	PCSD 128	EPA 200.8	188261		
70154624102	PCSD 130	EPA 200.8	188261		
70154624103	PCSD 131	EPA 200.8	188261		
70154624104	PCSD 132	EPA 200.8	188261		
70154624105	PCSD 133	EPA 200.8	188262		
70154624106	PCSD 134	EPA 200.8	188262		
70154624107	PCSD 135	EPA 200.8	188262		
70154624108	PCSD 136	EPA 200.8	188262		
70154624109	PCSD 137	EPA 200.8	188262		
70154624110	PCSD 138	EPA 200.8	188262		
70154624111	PCSD 139	EPA 200.8	188262		
70154624112	PCSD 140	EPA 200.8	188262		
70154624113	PCSD 141	EPA 200.8	188262		
70154624114	PCSD 142	EPA 200.8	188262		
70154624115	PCSD 143	EPA 200.8	188262		
70154624116	PCSD 144	EPA 200.8	188262		
70154624117	PCSD 145	EPA 200.8	188262		
70154624118	PCSD 147	EPA 200.8	188262		
70154624119	PCSD 150	EPA 200.8	188262		
70154624120	PCSD 151	EPA 200.8	188262		
70154624121	PCSD152 CAFE CLASS SINK 1 LEFT	EPA 200.8	188262		
70154624122	PCSD 153 - 2	EPA 200.8	188262		
70154624123	PCSD 154 - 3	EPA 200.8	188262		
70154624124	PCSD 155 - 4	EPA 200.8	188262		
70154624125	PCSD 156 - 5	EPA 200.8	188263		
70154624126	PCSD 157 CLASS SINK 6 RIGHT	EPA 200.8	188263		
70154624127	PCSD 158 RM 62 CLASS SINK	EPA 200.8	188263		
70154624128	PCSD 159 ACROSS 56 LOW DF BOT	EPA 200.8	188263		
70154624129	PCSD 160 2ND FLOOR STAFF BATH	EPA 200.8	188263		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



WO#: 70154624



70154624

# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

## Section C

### Required Project Information:

Company: Poland Central School District  
 Report To: Laura Dutton  
 Copy To:  
 Address: 74 Cold Brook Street  
 Poland, NY 13431  
 Email: lduion@polandcsd.org  
 Phone: 315-826-7900 Fax:  
 Project Name: school metal  
 Project #:

### Invoice Information:

Attention:  
 Company Name:  
 Address:  
 Pace Quote:  
 Pace Project Manager: julie.litvin@pacelabs.com,  
 Pace Profile #: x

ITEM #	MATRIX	CODE	COLLECTED		SAMPLE TYPE (G-GRAB C-COMP)	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
			START	END								
1	PCSD1	DW	11:20	0655	DW		11-22	07:30		11-23	0750	
2	PCSD2	DW	1	07:5	DW							
3	PCSD3	DW			DW							
4	PCSD4	DW		0659	DW							
5	PCSD5	DW		0700	DW							
6	PCSD6	DW		0701	DW							
7	PCSD7	DW		0656	DW							
8	PCSD8	DW		0657	DW							
9	PCSD9	DW		0658	DW							
10	PCSD10	DW		0702	DW							
11	PCSD11	DW		0747	DW							
12	PCSD12	DW		0711	DW							
ADDITIONAL COMMENTS: 5 Boxes + 120 box 120 box												
SAMPLE ID One Character per box. (A-Z, 0-9 /, -) Sample IDs must be unique												
MATRIX CODE DW Drinking Water WT Water WW Waste Water P Product SL Soil OL Oil WP Wipe AR Air OT Other TS Tissue												
PRESERVATIVES H2SO4 HNO3 HCl NaOH Na2S2O3 Methanol Other												
ANALYSES TEST 2008 Lead												
Residual Chlorine (Y/N)												

SAMPLER NAME AND SIGNATURE  
 PRINT Name of SAMPLER: *Chris Puz*  
 SIGNATURE of SAMPLER: *Chris Puz*  
 DATE Signed: 11-20-2022







# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company:	Poland Central School District	Report To:	Laura Dutton	Attention:	
Address:	74 Cold Brook Street	Copy To:		Company Name:	
	Poland, NY 13431			Address:	
Email:	ldutton@polandsd.org	Purchase Order #:		Pace Quote:	
Phone:	315-826-7900	Project Name:	school metal	Pace Project Manager:	julie.litvin@pacelabs.com
Requested Due Date:		Project #:		Pace Profile #:	x

ITEM #	MATRIX Drinking Water Water Waste Water Product Soil/Solid Oil Wipe Air Other Tissue	CODE DW WT WW P SL OL WP AR OT TS	COLLECTED		SAMPLE TYPE (G=GRAB C=COMP)	SAMPLE TEMP AT COLLECTION		# OF CONTAINERS	Preservatives										Analyses Test	Y/N	Requested Analysis Filtered (Y/N)										Residual Chlorine (Y/N)																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																								
			START	END		DATE	TIME		DATE	TIME	H2SO4	HNO3	HCl	NaOH	Na2S2O3	Methanol	Other																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																						
25					DW																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																		

ADDITIONAL COMMENTS	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
covered Printout Forming				Heidi-Heidi	11/20	1045	17.5 N NY
are not shot off							
but not in use							
SAMPLER NAME AND SIGNATURE				TEMP in C			
PRINT Name of SAMPLER: CLrb P-24							
SIGNATURE of SAMPLER: [Signature]				DATE Signed: 11-20-2010			
				Received on (Y/N)			
				Sealed (Y/N)			
				Cooler (Y/N)			
				Infect Samples (Y/N)			



# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company: Poland Central School District		Report To: Laura Dutton		Attention:	
Address: 74 Cold Brook Street		Copy To:		Company Name:	
Poland, NY 13431				Address:	
Email: ldutton@polandsd.org		Purchase Order #:		Pace Quote:	
Phone: 315-826-7900		Project Name: school metal		Pace Project Manager: julie.litvin@pacelabs.com	
Requested Due Date:		Project #:		Pace Profile #: x	
SAMPLE ID One Character per box. (A-Z, 0-9 / , -) Sample IDs must be unique		MATRIX CODE DW: Drinking Water WT: Waste Water WW: Wastewater P: Product SL: Solid OI: Oil V: Vape A: Air OT: Other TS: Tissue		SAMPLE TYPE (G=GRAB C=COMP)	
COLLECTED		START DATE TIME		END DATE TIME	
SAMPLE TEMP AT COLLECTION		DATE		TIME	
PRESERVATIVES		H2SO4		HNO3	
HCl		NaOH		Na2S2O3	
Methanol		Other		Y/N	
ANALYSES TEST		200.8 Lead		Y/N	
Residual Chlorine (Y/N)					
Requested Analysis Filtered (Y/N)					
TEMP in C		Received on		Ice (Y/N)	
Custody (Y/N)		Sealed Cooler (Y/N)		Samples Intact (Y/N)	
SAMPLER NAME AND SIGNATURE		DATE SIGNED: 11-23-20		DATE SIGNED: 11-23-20	
PRINT Name of SAMPLER: Chris Mader		DATE SIGNED: 11-23-20		DATE SIGNED: 11-23-20	
SIGNATURE of SAMPLER:		DATE SIGNED: 11-23-20		DATE SIGNED: 11-23-20	





# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company:	Poland Central School District	Report To:	Laura Dutton	Attention:	
Address:	74 Cold Brook Street	Copy To:		Company Name:	
	Poland, NY 13431	Purchase Order #:		Address:	
Email:	ldutton@polandcsd.org	Project Name:	school metal	Pace Quote:	
Phone:	315-826-7900	Fax:		Pace Project Manager:	julie.litvin@pacelabs.com
Requested Due Date:		Project #:		Pace Profile #:	x

Page : 6 Of 6

ITEM #	MATRIX	CODE	SAMPLE TYPE (G=GRAB C=COMP)	COLLECTED		DATE	TIME	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS	Received on	TEMP in C	Ice (Y/N)	Custody (Y/N)	Sealed Cooler (Y/N)	Samples Intact (Y/N)
				START	END															
61	PCSD61	g-15	DW	1-20																
62	PCSD62	g-15 - OFF	DW																	
63	PCSD63		DW	0547																
64	PCSD64		DW	0548																
65	PCSD65		DW	0558																
66	PCSD66		DW	0646																
67	PCSD67		DW	0607																
68	PCSD68		DW	0608																
69	PCSD69		DW	0601																
70	PCSD70		DW	0602																
71	PCSD71		DW	0603																
72	PCSD72		DW	0604																

SAMPLER NAME AND SIGNATURE		DATE Signed:	
PRINT Name of SAMPLER:			
SIGNATURE of SAMPLER:			

TEMP in C

Received on

Ice (Y/N)

Custody (Y/N)

Sealed Cooler (Y/N)

Samples Intact (Y/N)



The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Page: 7 Of 7[illegible][illegible]



# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A		Section B		Section C	
Required Client Information:		Required Project Information:		Invoice Information:	
Company:	Poland Central School District	Report To:	Laura Dutton	Attention:	
Address:	74 Cold Brook Street	Copy To:		Company Name:	
Poland, NY 13431		Purchase Order #:		Address:	
Email:	ldutton@polandcsd.org	Project Name:	school metal	Pace Project Manager:	julia.litvin@paceelabs.com
Phone:	315-826-7900	Fax:		Pace Profile #:	x
Requested Due Date:		Project #:			

Page : 8 Of 8

Regulatory Agency	
State / Location	
NY	

ITEM #	MATRIX	CODE	COLLECTED		SAMPLE TYPE (G=GRAB C=COMP)	MATRIX CODE (see valid codes to left)	SAMPLE TEMP AT COLLECTION		# OF CONTAINERS	Preservatives								Analyses Test	Y/N	Requested Analysis Filtered (Y/N)										Residual Chlorine (Y/N)																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																														
			Drinking Water Water Waste Water Product Soil/Solid Oil Wipe Air Other Tissue	DW WT WW P SL OL WP AR OT TS			START	END		DATE	TIME	DATE	TIME	H2SO4	HNO3	HCl	NaOH			Na2S2O3	Methanol	Other																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																						
85		PCSD85				DW					11-20	0716																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																

ADDITIONAL COMMENTS		RELINQUISHED BY / AFFILIATION		DATE		TIME		ACCEPTED BY / AFFILIATION		DATE		TIME		SAMPLE CONDITIONS	

## CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:										Section B Required Project Information:										Section C Invoice Information:										Page: 9 Of 9																																																																																																																																											
Company: Poland Central School District										Report To: Laura Dutton										Attention:										Regulatory Agency																																																																																																																																											
Address: 74 Cold Brook Street										Copy To:										Company Name:										State / Location																																																																																																																																											
Poland, NY 13431										Purchase Order #:										Address:										NY																																																																																																																																											
Email: ldutton@polandcsd.org										Project Name: school metal										Pace Project Manager: julie.litvin@pacelabs.com,										State / Location																																																																																																																																											
Phone: 315-826-7900										Fax:										Pace Profile #: x										NY																																																																																																																																											
Requested Due Date:										Project #:										Pace Profile #: x										NY																																																																																																																																											
Requested Analysis Filtered (Y/N)										Preservatives										Y/N										Requested Analysis Filtered (Y/N)																																																																																																																																											
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
ITEM #										MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)									
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH										Na2SO3										Methanol										Other										Analyses Test										200.8 Lead										Residual Chlorine (Y/N)																			
MATRIX CODE (see valid codes to left)										SAMPLE TYPE (G=GRAB C=COMP)										COLLECTED										SAMPLE TEMP AT COLLECTION										# OF CONTAINERS										Unpreserved										H2SO4										HNO3										HCl										NaOH																																																																															





# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company:	Poland Central School District	Report To:	Laura Dutton	Attention:	
Address:	74 Cold Brook Street	Copy To:		Company Name:	
Poland, NY 13431		Purchase Order #:		Address:	
Email:	ldutton@polandcsd.org	Project Name:	school metal	Pace Project Manager:	julie.litvin@paceclabs.com
Phone:	315-826-7900	Requested Due Date:		Pace Profile #:	x
				Regulatory Agency	
				State / Location	
				NY	

Page : 10 Of 10

ITEM #	MATRIX	CODE	COLLECTED		SAMPLE TYPE (G=GRAB C=COMP)	MATRIX CODE (see valid codes to left)	RELINQUISHED BY / AFFILIATION		DATE	TIME	ACCEPTED BY / AFFILIATION		DATE	TIME	SAMPLE CONDITIONS		Received on	Ice	Custody	Sealed	Cooler	Samples	Interact
109	PCSD109	DW	START	END	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE	TIME	DATE
110	PCSD110	DW	4-26	0751	0752																		
111	PCSD111	DW																					
112	PCSD112	DW																					
113	PCSD113	DW																					
114	PCSD114	DW																					
115	PCSD115	DW																					
116	PCSD116	DW																					
117	PCSD117	DW																					
118	PCSD118	DW																					
119	PCSD119	DW																					
120	PCSD120	DW																					

ADDITIONAL COMMENTS		SAMPLER NAME AND SIGNATURE	
		PRINT Name of SAMPLER: <i>Ar. J. Peter</i>	
		SIGNATURE of SAMPLER: <i>Ar. J. Peter</i>	
		DATE Signed: <i>11-20-2023</i>	
		TEMP in C	
		Received on	
		Ice	
		Custody	
		Sealed	
		Cooler	
		Samples	
		Interact	



# CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company:	Poland Central School District	Report To:	Laura Dutton	Attention:	
Address:	74 Cold Brook Street Poland, NY 13431	Copy To:		Company Name:	
Email:	ldutton@polandcsd.org	Purchase Order #:		Address:	
Phone:	315-426-7900	Project Name:	school metal	Pace Project Manager:	julie.litvin@pacelabs.com
Requested Due Date:		Project #:		Pace Profile #:	x

ITEM #	MATRIX	CODE	SAMPLE ID One Character per box. (A-Z, 0-9 /, -) Sample Ids must be unique	MATRIX CODE (see valid codes to left)		SAMPLE TYPE (G=GRAB C=COMP)		COLLECTED				SAMPLE TEMP AT COLLECTION	# OF CONTAINERS	Preservatives								Analyses Test	Y/N	200.8 Lead	Residual Chlorine (Y/N)																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
--------	--------	------	---	---------------------------------------	--	-----------------------------	--	-----------	--	--	--	---------------------------	-----------------	---------------	--	--	--	--	--	--	--	---------------	-----	------------	-------------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

ADDITIONAL COMMENTS		RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS	
								TEMP in C	Received on
								(Y/N)	Ice
								(Y/N)	Sealed
								(Y/N)	Cooler
								(Y/N)	Intact
								(Y/N)	Samples

Signature: *Julie Litvin* Date: 11-20-2018

Signature: *Chris Porel* Date: 11-20-2018

## CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

[illegible][illegible]

ADDITIONAL COMMENTS	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS		TEMP IN C	Received on	(Y/N)	Custody Sealed (Y/N)	Cooler (Y/N)	Samples Intact (Y/N)
				<i>H. O'Connell</i>	<i>11/20/2017</i>		<i>N N</i>							
<b>SAMPLER NAME AND SIGNATURE</b>														
<b>PRINT Name of SAMPLER:</b>														
<b>SIGNATURE of SAMPLER:</b>														

## CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

[illegible]



**CLIENT: PCSD**

## CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

[illegible]



# Sample Condition Upon Receipt

Client Name:

Project:

WO#: 70154624

PM: JL1

Due Date: 12/10/20

CLIENT: PCSD

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☐ Pace ☐ Other

Tracking #:

7721 6256 6923

Custody Seal on Cooler/Box Present: ☐ Yes ☐ No Seals intact: ☐ Yes ☐ No

Temperature Blank Present: ☐ Yes ☐ No

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☐ Ziploc ☐ None ☐ Other

Type of Ice: Wet Blue None

Thermometer Used: TH091

Correction Factor:

-0.2

☐ Samples on ice, cooling process has begun

Cooler Temperature (°C):

11.5

Cooler Temperature Corrected (°C):

11.3

Date/Time 5035A kits placed in freezer

Temp should be above freezing to 6.0°C

USDA Regulated Soil (☐ N/A, water sample)

Date and Initials of person examining contents:

HR 11/24/20

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? ☐ YES ☐ NO

Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? ☐ Yes ☐ No

If Yes to either question, fill out a Regulated Soil Checklist (F-LI-C-010) and include with SCUR/COC paperwork.

		COMMENTS:
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	11. Note if sediment is visible in the dissolved container.
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12. Sample "PCSD 6" marked absent but sample is present. Sample "PCSD 146" missing
-Includes date/time/ID/Analysis Matrix SL WT OIL		
All containers needing preservation have been checked	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO <sub>3</sub> <input type="checkbox"/> H <sub>2</sub> SO <sub>4</sub> <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot #	HC904495	
All containers needing preservation are found to be in compliance with EPA recommendation? (HNO <sub>3</sub> , H <sub>2</sub> SO <sub>4</sub> , HCl, NaOH>9 Sulfide, NaOH>12 Cyanide) Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sample #
		Initial when completed: Lot # of added preservative: Date/Time preservative added
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	14.
KI starch test strips Lot #		
Residual chlorine strips Lot #		Positive for Res. Chlorine? Y N
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	15.
Trip Blank Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	16.
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Pace Trip Blank Lot # (if applicable):		

Client Notification/ Resolution:

Field Data Required?

Y / N

Person Contacted:

Date/Time:

Comments/ Resolution: